


Instruction for FeLV Ag test



Sample fill line (black line) : 1 drop (10 μ l)

1

Sampling
& loading




X 2
(two times)

2 drops of sample

•• Do not overload the sample!!

2

Absorption




Wait for complete absorption

3

Buffer addition

00:00:00



2 drops of buffer
•• after complete absorption



4

Reading

00:05:00~00:10:00



+



Read the result between 5 ~ 10 min

-



5

Discard

00:10:00



Discard the used device
In valid result after 10 min

